



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## Certificate of Accreditation

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

### ***Superior Measuring LLC***

***1625 East Avis Drive, Madison Heights, MI 48071***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:*

### **ISO/IEC 17025:2005**

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated January 2009):

### ***CMM Dimensional Inspection*** ***(As detailed in the supplement)***

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

*Initial Accreditation Date:*

November 28, 2010

*Issue Date:*

July 8, 2016

*Expiration Date:*

September 30, 2018

*Accreditation No.:*

67476

*Certificate No.:*

L16-288

Tracy Szerszen  
President/Operations Manager

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: [www.pjllabs.com](http://www.pjllabs.com)*



# Certificate of Accreditation: Supplement

## Superior Measuring LLC

1625 East Avis Drive, Madison Heights, MI 48071  
Contact Name: John Bologna Phone: 248-616-4607

*Accreditation is granted to the facility to perform the following testing:*

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Dimensional Inspection <sup>F</sup>	Tools, Gages, and Manufactured Components	3 Dimensional Measurement with Coordinate Measurement Machine	Customer Supplied Dimensional Information & ANSI Y 14.5	X = 4 200 mm Y = 2 000 mm Z = 1 600 mm Detection limit 0.007 mm
		1 Dimensional Measurement with Micrometer	Customer Supplied Dimensional Information & ANSI Y 14.5	X = 26 mm Detection limit 0.005 mm

- The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer<sup>F</sup> would mean that the laboratory performs this testing at its fixed location.

